



July 15, 2014

Re: **Compatibility of Revital-Ox™ Resert® XL HLD High Level Disinfectant and Olympus Endoscopy Equipment**

Dear Health Care Professional:

This letter provides information on the material and functional compatibility of Olympus endoscopy equipment with STERIS Revital-Ox™ Resert® XL HLD High Level Disinfectant. Olympus has not tested the compatibility of Resert XL with Olympus endoscopy equipment. As a result, Olympus does not list Resert XL as either a compatible or incompatible product for the reprocessing of Olympus endoscopy equipment. For information on the efficacy of STERIS Resert XL, please contact STERIS Corporation (Mentor, OH).

The use of Resert XL does not by itself void the Olympus new product warranty for damage caused by faulty materials or workmanship. However, chemically induced damage (e.g., cracking of adhesives, loss of marking paint, insertion tube peeling, deterioration, floppiness or delamination, and unexplained fluid entry into the endoscope) resulting from the use of a liquid chemical germicide not listed by Olympus as compatible with Olympus flexible endoscopes will not be covered by warranty.

#### **NON-OLYMPUS REPAIRS**

Olympus cannot guarantee that any product repaired by a non-Olympus facility will be compatible with STERIS Resert XL. Moreover, if a product is repaired by a non-Olympus facility, the corresponding product warranty may be voided. In the event that your equipment has been repaired by such a facility, please contact the repair facility directly for compatibility information.

#### **WARRANTY**

Nothing contained in this letter alters, extends, or modifies in any way the authorized Olympus warranty applicable to each device or instrument.

If you have any questions, please contact your local Olympus sales representative or the Olympus Technical Assistance Center at 1-800-848-9024 (United States) or 1-800-387-0437 (Canada).

Sincerely,

Mary Ann Drosnock, MS, CIC, CFER, RM (NRCM)  
Manager, Infection Control  
maryann.drosnock@olympus.com

OLYMPUS AMERICA INC.  
3500 CORPORATE PARKWAY  
CENTER VALLEY, PA 18034  
TELEPHONE: (484) 896-5000  
www.olympusamerica.com

OLYMPUS CANADA  
25 Leek Crescent  
Richmond Hill, Ontario  
L4B 4B3, Canada  
Telephone: 1 (800) 387-0437  
www.olympuscanada.com